

National Institute of Cardiovascular Diseases

National (H.D) Shahhead Road, Karachi - Pakistan
Tel No: 3331-8330, 8830-1271 Ext. 419, Fax: 920-1289
www.nicvd.org

Dated: 20-07-2026

CORRIGENDUM

Tender for Surgical Headlights or Lead Lamps -08Nos - TENDER NO 51-2026


Reference to advertisements published in Daily Jang, Dawn & Awami Awaz, dated: 09-07-2026, regarding above tender. Below mentioned Technical Specification have been revised with the approval of competent authority.

Amended Technical Specification are as follows:

S.No	Item description.
01	Surgical Headlights or LED Lamps (08 Nos. Adult + 03 Nos. Peads) Technical Details: Best technology to accomplish the once impossible: a comfortable, cordless/with Cord. Head light that is versatile across many specialties in the clinic or operating room and Should enhance your surgical experience in terms of both intensity and spot size. • LED Light Sources with head light • Should Have 60,000 or more lux at 1 m working distance • Should have Intuitive user interface • System should have bright, evenly distributed light spot with headlights • Should have Battery service life (4-24 hours) • Headlight should be light weight maximum 250g on-word. • Should have USB charging time 4 hours for 2 batteries. • Should have Full beam angle adjustability and increased spot size range from 20mm on-word. • Should have Crisp, circular beam of homogenous light. • The head light should be Interchangeable headband for hard or soft top. • Color Temperature: 5500K - 6000K pure white Light. • LED Life: 45000 Hrs and Above. Warranty: 3 Years warranty with Parts. Certification: CE marked and ISO Certified. Country of Origin: Europe, USA, Japan <u>Note: Demonstration and Recommendation of End-user is Mandatory.</u>

NOTE:

Remaining other terms and conditions will remain the same as per tender notice.
If any vendor already participated in the tender in EPADS should re-submit your bid in EPADS.


EXECUTIVE DIRECTOR